

## DECLARATION OF COMPLIANCE

Pursuant to Act No. 22/1997 Coll., Section 13, Paragraph 2 in the wording of Act No. 71/2000 Coll. and Governmental Ordinance No. 181/2001 Coll., as amended

The manufacturer of **JORDAN** rehabilitation, gynaecologic and cosmetic reclining chairs

### **RESI Třeboň spol. s r.o.**

Specification of the types and varieties of **JORDAN** reclining chairs:

**A1, A2, A3, B2e, B3e, B5e, C1e, C2e, D, E1e, E2e, E1h, E2h, F2e, F3e, F5e, F2h, F3h, F5h, G3e, G5e, G5h, H3e, Gynaecologic chair, J6e, K8e, L10e, ME2e, ME3e, U3e, T3, M6e Osteo**

These **JORDAN** tables are designated for use in healthcare facilities for rehabilitation, physiotherapy, gynaecology and cosmetics.

#### **Information about the applied COMPLIANCE assessment method**

We used Section 12, Paragraph 3 Letter a) of Act No. 22/1997 Coll., as amended, on the technical requirements for products.

Assessment of the product was performed pursuant to Governmental Ordinance No. 181/2001 Coll., as amended. Classification pursuant to Section 6 of Governmental Ordinance No. 181/2001 Coll. (Class I)

The following technical regulations and harmonised technical standards were used to assess COMPLIANCE:



**ČSN EN 60601-1**

**ČSN EN 60601-2-38**

**TECHNICAL CONDITIONS No. VD 001/2002**

Information about the accredited entity participating in the assessment of compliance:

**Strojírenský zkušební ústav, s.p., Hudcova 56b, 621 00 Brno, Czech Republic, Identification Number: 00001490** issued the Final Report No. 31-9143/E dated 21 April 2000 and Certificate No. B-31-00453/00 dated 13 April 2000.

**We declare that the healthcare device fulfils the basic requirements of Governmental Ordinance No. 336/2004 Coll. It is safe for the designated use under the usual conditions, and is designated for repeated use under the stipulated hygienic conditions. Measures have been undertaken to ensure the compliance of all healthcare devices introduced to the market.**

Třeboň dated 9. Juli 2008

Ing. Jiří Šimeček